

**BAM**Bundesanstalt für
Materialforschung
und -prüfung12200 Berlin, Germany
T: +49 30 8104-0
F: +49 30 8104-7 2222

CERTIFICATE OF ANALYSIS

ERM®-FA001

POLYSTYRENE		
Molar mass M_w ¹⁾	Weight averaged molar mass M_w	
	Certified value ³⁾ [g mol ⁻¹]	Uncertainty ⁴⁾ [g mol ⁻¹]
	87600	2245
Viscosity ²⁾	Intrinsic viscosity [η]	
	Certified value ³⁾ [mL g ⁻¹]	Uncertainty ⁴⁾ [mL g ⁻¹]
	42.37	0.83
<p>1) obtained by laser light scattering 2) obtained by viscometry using an UBBELOHDE viscometer according to DIN 51562 – 1 3) Unweighted mean value of the means of accepted sets of data, each set being obtained in a different laboratory. The certified value is traceable to the International System of units (SI). 4) The certified uncertainty is the expanded uncertainty estimated in accordance with the Guide to the Expression of Uncertainty in Measurement (GUM) with a coverage factor $k = 2$, corresponding to a level of confidence of about 95 %. The certified uncertainty value is traceable to the International System of units (SI).</p>		

This certificate is valid for two years after purchase.

Sales date:

The minimum amount of sample to be used is 10 mg.

Accepted as an ERM®, Berlin,

Latest revision

Berlin, 2021-04-29

BAM Berlin
Department 6
Materials Chemistry
12200 Berlin, GermanyPriv. Doz. Dr. Franziska Emmerling
(Head of Department)BAM Berlin
Division 6.3
Structure Analysis
12200 Berlin, GermanyDr. Steffen Weidner
(Study Director)

CERTIFICATE

Indicative Values		
Molar masses ¹⁾	Weight-averaged molar mass M_w	
	Indicative value ²⁾ [g mol ⁻¹]	Uncertainty ³⁾ [g mol ⁻¹]
	79840	970
	Number-averaged molar mass M_n	
	Indicative value ²⁾ [g mol ⁻¹]	Uncertainty ³⁾ [g mol ⁻¹]
	71940	610
	Molar mass at peak maximum M_p	
	Indicative value ²⁾ [g mol ⁻¹]	Uncertainty ³⁾ [g mol ⁻¹]
	83650	1000

Additional Material Information	
Glass transition temperature T_g	104.1 °C ¹⁾
Density	1.04 g mL ⁻¹ ²⁾
Melt flow index	52 g 10 min ⁻¹
1) obtained by Differential Scanning Calorimetry (DSC) 2) at 25°C according to DIN EN ISO 1183-1 3) 5 kg at 200°C according to DIN EN ISO 1133	

NOTE

European Reference Material ERM®-FA001 was originally certified as BAM-P01. It was produced and certified under the responsibility of Bundesanstalt für Materialforschung und –prüfung (BAM) according to the principles laid down in the technical guidelines of the European Reference Materials® co-operation agreement between BAM-LGC-IRMM. Information on these guidelines is available on the Internet (<http://www.erm-crm.org>).

DESCRIPTION OF THE SAMPLE

The material was synthesised by different polymerization procedures and purified by dissolution and precipitation. A detailed homogeneity study was performed. The material identity was confirmed by Proton Nuclear Magnetic Resonance (NMR) and Infrared (IR) Spectroscopy. The sample consists of a white, amorphous material. It was bottled in glass vials with a unit size from 1 to 10 g.

ANALYTICAL METHOD USED FOR CERTIFICATION

- Size Exclusion Chromatography according to DIN 55 672 – 1 (GPC using tetrahydrofurane as eluent)
- Laser Light Scattering
- Viscometry according to DIN 51 562 – 1 (Viscometry: Determination of kinematic viscosity using a Ubbelohde – Viscometer, Part1: Design and realisation of measurements)

Details on the procedure for analysis can be found in the corresponding certification report.

PARTICIPANTS

Aventis, Frankfurt / M.
Bundesanstalt für Materialforschung und -prüfung, Berlin
Bayer AG, Uerdingen
Bayer AG, Leverkusen
Bundeskriminalamt, Wiesbaden
BMW, Dingolfing
Fraunhofer Institut für Angewandte Polymerforschung, Teltow
RWTH Aachen, Institut für Kunststoffverarbeitung
Institut für Lacke und Farben, Magdeburg
Institut für Polymerforschung, Dresden
Martin-Luther-Universität, Halle-Wittenberg
Max-Planck-Institut für Polymerforschung, Mainz
Polymer Standards Service GmbH, Mainz
Röhm GmbH, Darmstadt
RWTH Aachen, Institut für Textilchemie und Makromolekulare Chemie
Goldschmidt AG, Essen
Technische Universität Dresden
Universität Bayreuth
Universität Erlangen-Nürnberg
Universität Essen
Universität Freiburg
Universität Hamburg, Institut für Technische und Makromolekulare Chemie
Universität Leipzig
Johannes-Gutenberg-Universität Mainz, Institut für Makromolekulare Chemie
Johannes-Gutenberg-Universität Mainz, Institut für Physikalische Chemie
Universität Osnabrück
Universität Stuttgart, Institut für Technische Chemie
Universität Stuttgart, Institut für Textil- und Faserchemie
Universität Ulm
Universität - Gesamthochschule Siegen
Viscotek GmbH, Weingarten

SAFETY INFORMATION

Specific safety information are not known

INSTRUCTIONS FOR USE

Before withdrawing a sample the bottle has to reach room temperature. After use, the bottle has to be closed and stored at the recommended temperature.

STORAGE

Samples have to be stored at $5 \pm 2^\circ\text{C}$.

However, BAM cannot be held responsible for changes that happen during storage of the material at the customer's premises, especially of opened samples.

LEGAL NOTICE

Neither BAM, its contractors nor any person acting on their behalf:

- (a) make any warranty or representation, express or implied, that the use of any information, material, apparatus, method or process disclosed in this document does not infringe any privately owned intellectual property rights; or
- (b) assume any liability with respect to, or for damages resulting from, the use of any information, material, apparatus, method or process disclosed in this document save for loss or damage arising solely and directly from the negligence of BAM.

NOTE

A detailed technical report describing the production, characterisation as well as the analytical procedures applied and the treatment of the analytical data used to certify ERM®-FA001 is available on request from BAM.

Supply of Reference Materials by Bundesanstalt für Materialforschung und –prüfung:
Richard-Willstätter-Straße 11, 12489 Berlin, Germany
Phone: +49 30 8104 2061 – Fax: +49 30 8104 1117
e-mail: sales.crm@bam.de – internet: www.bam.de